

## APPENDIX C



Winnipeg Regional  
Health Authority  
Caring for Health

Office régional de la  
santé de Winnipeg  
À l'écoute de notre santé

WRHA Bariatric Equipment Pool  
Phone HSC Paging 787-2071 Pager 4009  
FAX # 787-3311  
Health Sciences Centre, Room GD033  
820 Sherbrook Street, Winnipeg, R3A 1R9

### WRHA Bariatric Equipment Pool Loan Form

**Item Requested:** (√)

- \*Bariatric Lift 454kg/1000lb capacity
- Bariatric Patient Sling     2XL     3XL
- Bariatric Stretcher 318kg/700lb 30" width
- Bariatric Stretcher 454kg/1000lb 37" width
- Bariatric Slider     1.5XL blue     2XL orange
- \*Wheelchair 386kg/850lb     28"width     30"width     32" width
- \*Shower Commode 320kg/705lb 30" width, w/integrated grab handles
- \*Shower Commode 454kg/1000lb 30" width

\* Wheelchairs and commode require an appropriate OT/PT consult before ordering to ensure the correct size is provided for the patient.

Name of Facility: \_\_\_\_\_ Patient PHIN # \_\_\_\_\_

Requested by: \_\_\_\_\_  
Name Phone Number Fax Number

Physiotherapy/Occupational Therapy Contact: \_\_\_\_\_  
Name Phone number

Staff Safety/OESH Contact: \_\_\_\_\_  
Name Phone number

Cost Centre Number \_\_\_\_\_ Cost Centre Name \_\_\_\_\_

Loan request authorized by Manager or delegate:

\_\_\_\_\_  
Print Signature Date

- **NOTE:** If equipment is damaged or lost, payment for the repairs/replacement will be the responsibility of the facility/department to which the equipment was initially loaned. By authorizing this request you authorize payment for such loss or damages.
- This equipment has been loaned to a specific patient within your facility. Do not transfer equipment to another patient or facility.
- Inform the WRHA Bariatric Equipment Pool (Ph#. 787-2071, request pager 4009) if there are any changes, problems or concerns with the equipment (Monday – Friday, 7 am – 3 pm)
- When finished, please contact the WRHA Bariatric Equipment Pool to arrange the return of the equipment.
- Equipment must be appropriately cleaned/disinfected prior to its return to the WRHA Bariatric Equipment Pool.

Loaned by: \_\_\_\_\_ Equipment KN#: \_\_\_\_\_  
(WRHA Bariatric Equipment Pool Staff) (If applicable)

Date returned: \_\_\_\_\_ Received by: \_\_\_\_\_

Return Condition/Concerns \_\_\_\_\_

**\*\* FAX COMPLETED FORM TO 787-3311 \*\***

## WRHA Bariatric Equipment Pool Equipment List



### Arjo Carmina Shower Commode

705 lb Weight capacity  
30" Seat width  
21" seat height  
20" seat depth  
20" seat back height (from seat)

#### \*\* Safety Note...

Not visible in the picture are the sides of the commode which work as grab bars. These are ideal for assisting with the patients' stability when getting on or off of the commode.



### Mapleleaf Wheelchairs Shower Commode

1000 lb Weight capacity  
30" Seat width  
21" seat height  
20" seat depth  
20" seat back height (from seat)  
Swing-up arm rests  
Removable, swing-away, elevating legrests



### **Hill-Rom P8000 Stretcher**

700 lb Weight capacity  
30" X 75" surface  
23"-35" surface height range  
5" thick comfort mattress  
It will fit through a 37" door opening  
Full range of positioning



### **Gendron Bariatric Trauma Stretcher**

1000 lb Weight capacity  
37" X 80" surface  
20"-28" Surface height range  
CPR release at head end  
Electric Operation of Head, Knee,  
Trend and Height (must be plugged in  
when not in use)  
It will fit through a 40" door opening



### **Arjo MaxiSlide**

Available in 1.5 XL and 2XI width  
There is no weight capacity on patient  
sliders  
Sliders should always be used in  
pairs.  
Sliders are single patient use only.



### **Gendron XL2000 Wheelchair**

850 lb Weight capacity  
Available in 28, 30 & 32" seat width  
20" seat depth  
17" seat height (from floor)  
Elevating removable legrests  
I.V. Pole, O2 tank holder  
2 attendant push bar



### **T.H.E. Medical Titan X Patient Lift**

1000 lb Weight capacity  
2 XL and 3XL slings available  
Built in battery and charger  
Lift from floor capable  
Must be plugged in when not in use

**\*\*Safety Note...**

This lift can be very difficult to manoeuvre

Updated June 30, 2010

Contact Information:

Health Sciences Centre Paging at 204-787-2071 Pager 4009

Mon – Fri 0700-1500 hrs